

**BAM**Bundesanstalt für
Materialforschung
und -prüfung

Test Report

on Type Testing of Pin-Index Cylinder Valves for Medical Gases
Type MYC-10B according to indicated parts of CGA V-9:2005

BAM reference	DG-07-138
Copy	1st copy of 3 copies
Customer	tekno valves 1, Chitpur Ghat Lane KOLKATA - 700 002 WB, INDIA
Date of order	7 April 2008
Receipt of order	7 April 2008 (by email)
Content of order	Testing of pin-index cylinder valves type MYC-10B according to indicated parts of CGA-V9:2005 partly
Receipt of drawings	Up to May 2009
Receipt of samples	21 May 2008, 7 samples for the impact tests 21 May 2008, 3 samples for oxygen pressure surge test 25 August 2008, 12 samples for testing the valve charac- teristics according to CGA-V9
Test period	May 2008 to May 2009
Test location	BAM Working Group "Pressure Equipment - Accessories" BAM-Building 44 BAM Working Group "Safe Handling of Oxygen" BAM-Building 41/40
Test procedure according to	Indicated parts of CGA-V9:2005

This report consists of sheet 1 to 5 and the annexes to the procedure Pressure Equip-
ment DG-07-138. The pressures indicated in the report are over-pressures in bar.

Test reports may only be published completely and verbatim, without any additions. Written consent must be obtained from BAM before any
changes or excerpts can be made. This consent may be withdrawn. The test results apply exclusively to the test materials examined by BAM.

TEST REPORT

